

Technical Data Sheet

Edition Date: Revision Date: Revised by: 2016-06-20 2016-09-09 Annette Roy, QA Coordinator

Р	roduct Name: Micrewtube [®]	Catalogue No.: T341-5TPRTP
This document replaces any previous version		
1.	Product Description:	
	> 1.5ml Tamper Evident Micrewtube [®] :	Non sterile, conical bottom, disposable (tube only). These tubes are specially made to be used with

These tubes are specially made to be used with tamperproof caps and have printed graduations and a white marking area.

2. Packaging:

Case: 1 000 units per package

3. Product Specifications:

- Material: Polypropylene tube;
- Temperature range: -196°C to +121°C;
- Tubes have printed graduations and a white marking area;
- Certified RNase, DNase, Pyrogen and DNA Free;
- Centrifuged up to 20 000 g.

4. Standards and Conformity:

- ISO 2859-1: Sampling and inspection procedures
- FDA:Resin conforms to FDA 21 CFR 177.1520
- USP: Resin conforms to USP Class VI
- CONEG:

- Plastics and colorants are in conformity with CONEG standards for heavy metals
- REACH (SVHC): Plastic is in conformity to REACH standards
- > LATEX: Material is Latex free
- **BSE / TSE:** Material is BSE / TSE free

5. Quality Assurance:

- Clear, no presence of contamination in plastic;
- Visual attributes;
- Volume measurements;
- Closure verification;
- Leak proof testing in vacuum at 71.3cm Hg;
- Gas phase of Liquid Nitrogen resistance.



Technical Data Sheet

Edition Date: Revision Date: Revised by: 2016-06-20 2016-09-09 Annette Roy, QA Coordinator

6. Traceability:

Lot No. Composition: 8 to 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label;
- 2. On label inserted inside the master case;
- 3. On the inner bag.

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions;
- Avoid temperature variations and humidity;
- Protect from any possible contamination;
- Protect from any damage to the packaging.

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents;
- > Follow chemical resistance chart recommendations;
- > For use in automated equipment, follow the equipment manufacturer's instructions;
- > Should be used only in the gas phase of Liquid Nitrogen.

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.