



Brook Hill, Woodstock, Oxfordshire
 OX20 1TU, United Kingdom
 T: +44 (0)1993 812021
 F: +44 (0)1993 813466
 E: info@owenmumford.co.uk
 www.owenmumford.com

Authorised Representative

Owen Mumford GmbH
 Alte Hage 1
 63762 Groostheim
 Telefon: 06026/9775-0
 Telefax: 06026/9775-29
 www.owen-mumford.de
 E-Mail: info@owen-mumford.de

Sparkasse Miltenberg-Obernburg
 IBAN: DE 2879650000500201983
 SWIFT-BIC: BYLADEM1MIL

IK: 590 961 092

Amtsgericht Aschaffenburg, HRB 9177
 Geschaftsfuhrer: Adam J. Mumford, Thomas Robach

USt.-Id.-Nr.: DE 157922203
 Steuer-Nr. 204/134/90031

EC MEDICAL DEVICE DIRECTIVE DECLARATION OF CONFORMITY

This declaration is made in accordance the essential requirements of the Council Directive 93/42/EEC.

The device is classified under Annex IX of the directive as follows:

- | | |
|---------------------------------------|---------------------------------------|
| • Classification: | Class IIa |
| • Term: | Transient Use |
| • Invasive: | Invasive |
| • Rule: | 6 |
| • Conformity Assessment Route: | Annex II (excluding section 4) |

GMDN Code and Definition 61578 - A sterile, hand-held manual instrument intended to be used for controlled skin puncture/cut to obtain a capillary blood specimen, performed by a healthcare provider (e.g., on a neonate) or a patient (e.g., a diabetic), typically at the fingertip or ear lobe. It includes a preloaded lancet tip, and has a manually powered mechanism (e.g., spring-loaded) which enables the tip to puncture to a predetermined depth and blood subsequently to be squeezed out of the puncture site. This is a single-use device.

Brand Names

Code	Product Name	Product Specification
000361	Unistik® Touch 30G Lancet (Translucent) (100)	30 Gauge – 1.5 mm Lancing Depth
000362	Unistik® Touch 30G Lancet (Translucent) (200)	30 Gauge – 1.5 mm Lancing Depth
000363	Unistik® Touch 28G Lancet (Translucent) (100)	28 Gauge – 1.8 mm Lancing Depth
000364	Unistik® Touch 28G Lancet (Translucent) (200)	28 Gauge – 1.8 mm Lancing Depth
000365	Unistik® Touch 23G Lancet (Translucent) (100)	23 Gauge – 2.0 mm Lancing Depth
000366	Unistik® Touch 23G Lancet (Translucent) (200)	23 Gauge – 2.0 mm Lancing Depth
000367	Unistik® Touch 21G Lancet (Translucent) (100)	21 Gauge – 2.0 mm Lancing Depth
000368	Unistik® Touch 21G Lancet (Translucent) (200)	21 Gauge – 2.0 mm Lancing Depth
000874	Unistik® Touch 30G Lancet (Translucent) (5)	30 Gauge – 1.5 mm Lancing Depth
000875	Unistik® Touch 28G Lancet (Translucent) (5)	28 Gauge – 1.8 mm Lancing Depth
000876	Unistik® Touch 23G Lancet (Translucent) (5)	23 Gauge – 2.0 mm Lancing Depth

This information is confidential to Owen Mumford. The user is responsible for confirming use of the current approved revision prior to use.

The above device has been designed and manufactured within a quality system established and maintained by Owen Mumford Limited. The quality system, along with the relevant technical documentation is periodically reviewed by SGS Belgium NV; SGS House Noorderlaan 87 2030, Antwerp, Belgium. Notified Body number CE 1639.

Name: **Darren Mansell**

Signature: 

Job Title: **Regulatory Affairs Manager**

Date: **01 DEC 2020**

**For Mr. J. Severn
Managing Director – Owen Mumford Ltd.**