

Declaration of Conformity

| | | | | |
|-------------------------------------|---|----|-----|--|
| Product Name | Histology Cassettes | | | |
| Intended use | Hold tissue specimens for further in vitro diagnostic | | | |
| Product Codes | M385; M386; M392; M405; M406; M407; M409; M460; M470; M471; M474; M475; M476; M478; M480; M482; M483; M485; M486; M490; M491; M492; M493; M498; M499; M502; M503; M505; M506; M507; M508; M509; M510; M511; M512; M515; M516; M517; M518; M520; M521; M525; M526; M430; M431; M432; M433; M434; M435; M530; M531 M532 M533 M534; M535 | | | |
| Basic UDI-DI | 667243 | | | |
| Manufacturer | Simport Scientific Inc. 2588 Bernard-Pilon Beloeil Quebec J3G 4S5 Canada | | | |
| Authorized Representative in Europe | <table border="1"> <tr> <td>EC</td> <td>REP</td> </tr> </table> | EC | REP | EMERGO EUROPE Westervoortsedijk 60 6827 AT Arnhem The Netherlands SRN NL-AR-000000116 |
| EC | REP | | | |

Simport declares that the above-mentioned products meet the provision of the Regulation (EU) 2017/746 of the European Parliament and of the Council on *In Vitro* Diagnostic Medical Devices and Regulation (EU) 2017/746 as transposed in the national laws of the Member States:

- That the products have been classified as general IVD and Rule 5(c), class A (Low Individual Risk and Low Public Health Risk);
- That the products listed above are in conformity with the Annex II (Essential Requirements) and III of Regulation (EU) 2017/746
- That the products do not contain medicinal substances;
- That the products do not contain animal tissues.



Annette Roy, Regulatory and Technical Support
Place and Date: Beloeil, February 02, 2023