

Product Data (PD)

Universal Container PP 35 ml w/Screw Cap				
REF	08-323-1001	08-323-1000	08-323-3001	

1.	Description / Specification					
1.01	Description	Universal Container 35 ml w/Screw Cap				
1.02	Dimension	See Customer Drawing				
1.03	Volume	35 ml				
1.04	Material / Resin	Cup	Polypropylene (PP)			
		Lid	Polyethylene (HDPE)			
	Properties of Resin	See www.boettger.com				
1.05	Colour	Cup Lid	Translucent			
			08-323-1001	08-323-1000	08-323-3001	
			Blue	Natural	Blue, mounted	
1.06	Hygiene Level	Standard hygienic		Irradiation 10 kGy <table border="1" data-bbox="1157 963 1316 1019"><tr><td>10 kGy</td><td>β</td></tr></table>	10 kGy	β
10 kGy	β					
1.07	Quality Control	Testing of attributive and variable characteristics in accordance with the valid specification				
1.08	Storage	To be stored in closed original packaging: Dry, at room temperature				
1.09	Usage / Purpose	General laboratory product, Collection and evaluation of samples				
1.10	Conformities / Regulations	See Classification				
1.11	Product Stewardship	See Product Stewardship				
1.12	Other information	For single use only				
2.	Features					
2.01	Basic features	-				
2.02	Temperature range	Cup	-196°C to 121°C			
		Lid	-50°C to 70°C			
2.03	Autoclavability	Cup	Yes			
		Lid	No			
2.04	Centrifugation	N/A				
2.05	Chemical Resistance	See www.boettger.com				
2.06	Shelf life / Intended Period of Usage	N/A	3 years from LOT			
		3 years from LOT	N/A			
2.07	Other information	Suitable for Formalin				
3.	Packaging	08-323-1001	08-323-1000	08-323-3001		
3.01	Pcs/Bag	500	500	500		
3.02	Pcs/Ctn	500	500	500		
3.03	Pcs/Pallet	12000	12000	12000		
3.04	LOT / VP	WWYY (week, year of production) / XX (packaging)				
3.05	Other information	See Measures				
4.	Other Information					
4.01	Codes and Numbers	See Codes and Numbers				
4.02	Symbols	See Symbols				

All informations without guarantee. Above mentioned products are manufactured according to company specifications and our certified Quality Management System corresponding to DIN EN ISO 13485. The quality check is carried out on a sample basis; an individual examination is not possible. Deficiencies parts may occur. We accept no liability for consequential damages.

Revision	Drawn	Approved	Confidential: Information contained in this document or drawing is confidential and proprietary to Paul Boettger GmbH & Co. KG. This document may not be reproduced for any reason without written permission from Paul Boettger GmbH & Co. KG. All rights of design, invention and copyright are reserved.
06	22.05.2019 QMB	22.05.2019 QMB	