

Product Data (PD)

Universal Container PP 20 ml w/Screw Cap REF | 08-313-1001 | 08-313-1000 | 08-313-3001 |

1.	Description / Specification			
1.01	Description	Universal Container PP 20 ml w/Screw Cap		
1.02	Dimension	See Customer Drawing		
1.03	Volume	20 ml		
1.04	Material / Resin	Cup	Polypropylene (PP)	
		Сар	Polyethylene (HDPE	-)
	Properties of Resin	See www.boettger.c		
1.05	Colour Cup	Translucent		
	Lid	08-313-1001	08-313-1000	08-313-3001
		Blue	Natural	Blue, mounted
1.06	Hygiene Level	Standard hygienic		Irradiation 10 kGy
				10 kGy β
				10 kGy β
1.07	Quality Control	Testing of attributive	and variable characte	eristics in accordance
		with the valid specifi		
1.08	Storage	To be stored in close	ed original packaging:	
	_	Dry, at room temper		
1.09	Usage / Purpose	General laboratory product,		
		Collection and evalu	ation of samples	
1.10	Conformities / Regulations	See Classification		
1.11	Product Stewardship	See Product Stewardship		
1.12	Other information	For single use only		
2.	Features			
2.01	Basic features	-	T	
2.02	Temperature range	Cup -196°C to 121°C		
0.00		Lid -50°C to 70°C		
2.03	Autoclavability	Cup	Yes	
0.04	0	Lid	No	
2.04	Centrifugation	N/A		
2.05	Chemical Resistance	See <u>www.boettger.com</u> N/A 3 years from LOT		
2.06	Shelf life / Intended Period of Usage	-		3 years from LOT N/A
2.07	Other information	3 years from LOT N/A Suitable for Formalin		
3.	Packaging	08-313-1001	08-313-1000	08-313-3001
3.01	Pcs/Bag	500	500	500
3.02	Pcs/Ctn	1000	1000	1000
3.03	Pcs/Pallet	20000	20000	20000
3.04	LOT / VP			
3.05	Other information	WWYY (week, year of production) / XX (packaging) See <i>Measures</i>		
4.	Other Information			
4.01	Codes and Numbers	See Codes and Numbers		
4.02	Symbols	See Symbols		

All informations without guarantee. Above mentioned products are manufactured according to company specifications and our certified Quality Management System corresponding to DIN EN ISO 13485. The quality check is carried out on a sample basis; an individual examination is not possible. Deficiencies parts may occur. We accept no liability for consequential damages.

Revision	Drawn	Approved	Confidential: Information contained in this document or drawing is confidential
07	18.09.2019	18.09.2019	and proprietory to Paul Boettger GmbH & Co. KG. This document may not be
07	QMB	QMB	reproduced for any reason without written permission from Paul Boettger GmbH & Co. KG. All rights of design, invention and copyright are reserved.