



TECHNICAL DATA SHEET

PRODUCT DESCRIPTION

PRODUCT CODE	141C.USE
GENERAL DESCRIPTION	TRANS STUART LQ A/PL
DETAILED DESCRIPTION	1 box containing 8 Vi-Pak, each containing 50 film-film pouches each containing 1 tube with sponge and 1.2 ml of medium with liquid Stuart with white cap without receptacle and a plastic applicator with white cap

PRODUCT COMPONENTS

MEDIUM	COMPOSITION
STUART LIQUID	Sodium Glycerophosphate Calcium Chloride Mercaptoacetic Acid Distilled water

		SPECIFICATIONS
TRANSPORT TUBE	TUBE	TUBE LENGTH: 154.50 mm ± 1.00 mm INSERTION DIAMETER: 9.00 mm ± 0.50 mm EXTERNAL DIAMETER: 12.00 mm ± 0.50 mm MATERIAL: POLYPROPYLENE

This bulletin (Rev.06-08/05/2023) is for your guidance and is based on information believed to be reliable and should be periodically reviewed.



TECHNICAL DATA SHEET

SPONGE	SPONGE	SPECIFICATIONS
		LENGTH: 37.50 mm + 2.00 mm -1.50 mm WIDTH: 10.00 mm ± 0.50 mm THICKNESS: 5.00 mm ± 0.50 mm MATERIAL: POLYURETHANE
CAP	WHITE CAP W/O RECEPTACLE	SPECIFICATIONS
		EXTERNAL DIAMETER: 12.00 mm ± 0.10 mm CAP LENGTH: 32.80 mm ± 0.20 mm MATERIAL: HIGH DENSITY POLYETHYLENE
APPLICATOR	PLASTIC APPLICATOR	SPECIFICATIONS
		TOTAL LENGTH: 147.00 mm ± 2.00 mm STICK DIAMETER: 2.50 mm ± 0.10 mm FIBER TIP LENGTH: 16.00 mm ± 3.00 mm FIBER TIP DIAMETER: 4.50 mm ± 0.50 mm STICK MATERIAL: POLYSTYRENE FIBER TIP MATERIAL: RAYON

This bulletin (Rev.06-08/05/2023) is for your guidance and is based on information believed to be reliable and should be periodically reviewed.



TECHNICAL DATA SHEET

CAP	WHITE CAP	SPECIFICATIONS
		EXTERNAL CAP DIAMETER: 12.00 mm \pm 0.10 mm CAP LENGTH: 32.80 mm \pm 0.20 mm INSERTION DIAMETER: 9.30 mm \pm 0.05 mm MATERIAL: HIGH DENSITY POLYETHYLENE

PACKAGING SPECIFICATION

OUTER BOX QTY	400
PACKAGING	FILM-FILM POUCH
VI-PAK QUANTITY	50

IMPORTANT INFO

STERILIZATION METHOD	IONIZING IRRADIATION
STORAGE TEMPERATURE	5/25°C-DO NOT FREEZE
SHELF LIFE	15 MONTHS
MEDIUM pH	7.2 \pm 0.3

This bulletin (Rev.06-08/05/2023) is for your guidance and is based on information believed to be reliable and should be periodically reviewed.