



Resch Laborglas Wechterswinkler Str 7 97618 Unsleben

Datenblatt

Artikel	Pasteurpipette
Artikelnummer	00410230
Material	AR Glas Hydrolytic Class (DIN ISO 719) HGB 3 Acid Class (DIN 12116) Class S 1 Alkali Class (DIN ISO 695) Class A 2 Freigabe European Pharmacopoeia 3.1.3 Polyolefines (6 Edition -2008) 3.1.4 Polyethylene without Additives for Containers for Preparations for Parenteral Use and for Ophthalmic Preparations 8 6 Edition - 2008)
Dimension	Länge 225 + 5 mm (Abmessungen nach ISO 7712)