

Datenblatt - *Data Sheet*

Kunde
Customer:

MLS nv
Ringlaan 7
B-8930 Menen

Artikel
Article:

**Test Tube 75x12x0,5-0,6 Boro 5.1; round bottom; rim
melted**

Artikelnr.
Article no.:

A507512000611

Material:
Material:

Glas / glass: Borosilicate 5.1

Water resistance tested to ISO 719 (at 98 °C): HGB 1
Alkali resistance tested to ISO 695: class A2
Acid resistance tested to ISO 12116: S1

Abmessungen/Norm:
Dimension/Standard:

Acc. ISO 4142

Height H: **75 mm**
Outside diameter D: **12 mm**
Wall Thickness W: **0,5-0,6 mm**

