



TECHNICAL DATA SHEET

PRODUCT DESCRIPTION

PRODUCT CODE	949C
GENERAL DESCRIPTION	SRK 10 ML UNS BULK 100 MM TUBE
DETAILED DESCRIPTION	1 outer box contain 6 inner boxes each composed of 50 tubes with green cap, filled with 10mL of UNS medium, each containing 1 rayon tip swab.

PRODUCT COMPONENTS

MEDIUM	COMPOSITION												
UNS MLS	<table> <tr> <td>Peptone</td> <td>Sodium thiosulfate</td> </tr> <tr> <td>Sodium Chloride</td> <td>Sodium Thioglicollate</td> </tr> <tr> <td>Disodium phosphate</td> <td>Sodium Bicarbonate</td> </tr> <tr> <td>Monopotassium phosphate</td> <td>Potassium chloride</td> </tr> <tr> <td>Tween 80</td> <td>Osmotic water</td> </tr> <tr> <td>Lecithin</td> <td></td> </tr> </table>	Peptone	Sodium thiosulfate	Sodium Chloride	Sodium Thioglicollate	Disodium phosphate	Sodium Bicarbonate	Monopotassium phosphate	Potassium chloride	Tween 80	Osmotic water	Lecithin	
Peptone	Sodium thiosulfate												
Sodium Chloride	Sodium Thioglicollate												
Disodium phosphate	Sodium Bicarbonate												
Monopotassium phosphate	Potassium chloride												
Tween 80	Osmotic water												
Lecithin													

SAMPLING TUBE	TUBE WITH ROUNDED BOTTOM	SPECIFICATIONS
		TUBE LENGTH: 101.80 mm ± 0.20 mm TUBE DIAMETER: 16.10 mm ± 0.20 mm INTERNAL DIAMETER TUBE: 14.10 mm ± 0.20 mm MATERIAL: POLYPROPYLENE
CAP	GREEN CAP	SPECIFICATIONS
		LOWER CAP DIAMETER: 21.50 mm ± 0.20 mm CAP LENGTH: 16.30 mm ± 0.20 mm HIGHER CAP DIAMETER: 20.50 mm ± 0.20 mm MATERIAL: HIGH DENSITY POLYETHYLENE

This bulletin (Rev.01-14/12/2020) is for your guidance and is based on information believed to be reliable and should be periodically reviewed.



TECHNICAL DATA SHEET

APPLICATOR	PLASTIC APPLICATOR	SPECIFICATIONS
		TOTAL LENGTH: 85.50 mm \pm 2.00 mm STICK DIAMETER: 2.50 mm \pm 0.10 mm FIBER TIP LENGTH: 16.00 mm \pm 3.00 mm FIBER TIP DIAMETER: 4.50 mm \pm 0.50 mm STICK MATERIAL: POLYSTYRENE

PACKAGING SPECIFICATION

INNER BOX QTY	50
OUTER BOX QTY	300
PACKAGING	BULK

IMPORTANT INFO

STORAGE TEMPERATURE	5/25°C-DO NOT FREEZE
SHELF LIFE	12 MONTHS
MEDIUM pH	7.3 \pm 0.5

This bulletin (Rev.01-14/12/2020) is for your guidance and is based on information believed to be reliable and should be periodically reviewed.