

Product Data (PD)

Universal Container PP 100 ml w/Screw Cap				
REF	08-343-1001	08-343-1000	08-343-3001	

1.	Description / Specification					
1.01	Description	Universal Container PP 100 ml w/Screw Cap				
1.02	Dimension	See Customer Drawing				
1.03	Volume	100 ml				
1.04	Material / Resin	Cup	Polypropylene (PP)			
		Lid	Polyethylene (HDPE)			
	Properties of Resin	See www.boettger.com				
1.05	Colour	Cup Lid	Translucent			
			08-343-1001	08-343-1000	08-343-3001	
			Blue	Natural	Blue, mounted	
1.06	Hygiene Level	Standard hygienic		Irradiation 10 kGy <table border="1" data-bbox="1161 965 1318 1016"><tr><td>10 kGy</td><td>β</td></tr></table>	10 kGy	β
10 kGy	β					
1.07	Quality Control	Testing of attributive and variable characteristics in accordance with the valid specification				
1.08	Storage	To be stored in closed original packaging: Dry, at room temperature				
1.09	Usage / Purpose	General laboratory product, Collection and evaluation of samples				
1.10	Conformities / Regulations	See Classification				
1.11	Product Stewardship	See Product Stewardship				
1.12	Other information	For single use only				
2.	Features					
2.01	Basic features	-				
2.02	Temperature range	Cup	-196°C to 121°C			
		Lid	-50°C to 70°C			
2.03	Autoclavability	Cup	Yes			
		Lid	No			
2.04	Centrifugation	N/A				
2.05	Chemical Resistance	See www.boettger.com				
2.06	Shelf life / Intended Period of Usage	N/A		3 years from LOT		
		3 years from LOT		N/A		
2.07	Other information	Suitable for Formalin				
3.	Packaging	08-343-1001	08-343-1000	08-343-3001		
3.01	Pcs/Bag	250	250	250		
3.02	Pcs/Ctn	250	250	250		
3.03	Pcs/Pallet	5000	5000	5000		
3.04	LOT / VP	WWYY (week, year of production) / XX (packaging)				
3.05	Other information	See Measures				
4.	Other Information					
4.01	Codes and Numbers	See Codes and Numbers				
4.02	Symbols	See Symbols				

All informations without guarantee. Above mentioned products are manufactured according to company specifications and our certified Quality Management System corresponding to DIN EN ISO 13485. The quality check is carried out on a sample basis; an individual examination is not possible. Deficiencies parts may occur. We accept no liability for consequential damages.

Revision	Drawn	Approved	Confidential:
09	18.09.2019	18.09.2019	Information contained in this document or drawing is confidential and proprietary to Paul Boettger GmbH & Co. KG. This document may not be reproduced for any reason without written permission from Paul Boettger GmbH & Co. KG. All rights of design, invention and copyright are reserved.
	QMB	QMB	